

SEPTOSTERILE

High-level disinfectant for medical devices, heat-sensitive dental & endoscopic instruments


HISEPT SEPTOSTERILE is a high-level disinfectant for reusable medical devices, patient-care equipment, heat-sensitive dental and endoscopic instruments. This product contains glutaraldehyde 2.4% as a disinfection agent, with fast-acting and broad-spectrum antimicrobial activity. It eliminates virtually all recognized pathogenic microorganisms on medical devices, including all steam and heat-sensitive semi-critical instruments.

Characteristics

- Contains 2.4% glutaraldehyde
- Broad-spectrum antimicrobial activity
- Ready-to-use without the need of any activator
- Reusable for up to 28 days
- Limited evaporation of glutaraldehyde

Product Specification

Color/Form	Clear green solution
pH	5.5 - 6.6
Scent	Characteristic odor of aldehyde and lemon scent
Shelf Life	24 months



THIS IS A HISEPT PRODUCT, WITH
**HIGH-LEVEL
SAFETY &
PROTECTION**

Ingredients

100 g HISEPT SEPTOSTERILE contains:

- 2.4 g glutaraldehyde

Auxiliary ingredients:

- Surfactant
- Buffering agent
- Corrosion inhibitors
- Essence
- Dye
- Deionized water

Microbiological Effectiveness

Meets European Standards

Bactericidal Activity

According to EN 13727 (contact time: 10 minutes) Staphylococcus. aureus/ ATCC 6538
Enterococcus. hirae/ ATCC 10541
Pseudomonas. aeruginosa/ ATCC 15442
Escherichia. coli/ ATCC 10536

According to EN 14561 (contact time: 10 minutes) Staphylococcus. aureus/ ATCC 6538
Enterococcus. hirae/ ATCC 10541
Pseudomonas. aeruginosa/ ATCC 15442

Yeasticidal Activity

According to EN 13624 (contact time: 10 minutes) Candida. albicans/ ATCC 10231

According to EN 14562 (contact time: 10 minutes) Candida. albicans/ ATCC 10231

Fungicidal Activity

According to EN 13624 (contact time: 10 minutes) Aspergillus. niger/ ATCC 16404

According to EN 14562 (contact time: 30 minutes) Aspergillus. niger/ ATCC 16404

Sporicidal Activity

According to EN 13704 (contact time: 60 minutes) Bacillus. subtilis/ ATCC 6633

Virucidal Activity

According to EN 14476 (contact time: 10 minutes) Herpes simplex virus(HSV-1)

Directions for use

Cleaning process: Visible dirt must be thoroughly cleaned from medical devices before disinfection with HISEPT SEPTOSTERILE. Refer to the reusable device manufacturer's labeling for additional instructions on disassembly, decontamination, cleaning and leak testing of the equipment.

Disinfection: Immerse clean device completely into the ready-to-use HISEPT SEPTOSTERILE, filling all lumens and eliminating bubbles and cover the tank for at least 10 minutes. Remove device from the solution and rinse thoroughly with clean water.

Compatibility

HISEPT SEPTOSTERILE is compatible with most materials such as stainless steel, copper, plastic, rubber and silicon.

2/2 Latest revision: Dec 1, 2023


Packagings

3 canisters per carton, each 5 L

6 bottles per carton, 1 L each

Hazard Rating

Ready-to-use	NFPA
Health	2
Flammability	0
Reactivity	0



Warning

- The disinfection tank shall always be covered with a lid. If the solution is kept in a closed container, it will be reusable for up to 28 days.
- In case of contact with skin and eyes, rinse immediately with plenty of water and seek medical advice.
- Keep out of reach of children and unaware persons.
- Store between + 5 °C and + 35 °C and away from direct sunlight.



Certifications

- >> ISO 9001: 2015
- >> ISO 13485: 2016
- >> GMP
- >> FSC



HISEPT[®]
HIGH-LEVEL SAFETY & PROTECTION

Middle East Office
Behban Pharmed Lotus Co.
Address : Unit 113, 12 Floor, Bldg. B, Kaveh
Complex, Azadi Ave., Tehran, Iran
Tel: (+9821) 66581294-5/66581269

info@hisept.com

World Locations: Düsseldorf, Kuala Lumpur, Istanbul, Tehran, Baghdad